

Certificate of Analysis

Product Description: ToRCH Negative Plasma
Material No(s): 0320-0011, 0320-0012, 0320-0013, 0320-0014
Lot No: 10055497
Donor ID: BD201434
Gender: Female
Age: 42
Country of Origin: US
Collection Date: 02-Apr-14
Anticoagulant: 4% Sodium Citrate
Storage Temperature: -20 degrees C
Shipping Conditions: Dry ice
Intended Use: For research use only. Not for use in diagnostic procedures.
Safety Precautions: Use universal precautions when handling this material.
Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
20-Jul-15	CMV IgG	Trinity Biotech Captia EIA	0.019	S/CO	
20-Jul-15	HSV-1/2 IgG	Trinity Biotech Captia EIA	0.018	S/CO	
20-Jul-15	Rubella IgG	Trinity Biotech Captia EIA	0.011	S/CO	
20-Jul-15	Toxoplasma IgG	Trinity Biotech Captia EIA	0.002	S/CO	
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIV 1 RNA, HCV RNA, HBV DNA and found to be negative.					

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Labonte

19-Oct-15

Prepared By

Date

Christine Pawia

20-Oct-15

QA Verified By

Date

