

Certificate of Analysis

Product Description: Syphilis (T. Pallidum) IgM Positive Plasma
 Material No(s): 0365-0009, 0365-0010, 0365-0011, 0365-0012
 Lot No: 10118414
 Donor ID: BD245067
 Gender: Male
 Age: 30
 Country of Origin: NA
 Collection Date: 06-Dec-13
 Anticoagulant: NA
 Storage Temperature: -20
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
18-Jun-15	Syphilis IgM	Calbiotech EIA	2.667	S/CO	NA
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
18-Jun-15	Rapid Plasma Reagin (Syphilis)	Becton Dickinson RPR	POS	NA	Neat
18-Jun-15	Rapid Plasma Reagin (Syphilis)	Becton Dickinson RPR	POS	NA	1:2 dilution
18-Jun-15	Rapid Plasma Reagin (Syphilis)	Becton Dickinson RPR	POS	NA	1:4 dilution
18-Jun-15	Rapid Plasma Reagin (Syphilis)	Becton Dickinson RPR	NEG	NA	1:8 dilution
18-Jun-15	Rapid Plasma Reagin (Syphilis)	Becton Dickinson RPR	NEG	NA	1:16 dilution
18-Jun-15	Syphilis IgG	Trinity Biotech Captia EIA	2.051	S/CO	NA

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Labonte

15-Dec-15

Prepared By

Date

Tsering Chelung

16-Dec-15

QA Verified By

Date