

## Certificate of Analysis

|                      |  |
|----------------------|--|
| Product Description: | HCV Ag Positive Plasma                                       |
| Material No(s):      | 0310-0027, 0310-0028, 0310-0029                              |
| Lot No:              | 9174620  |
| Donor ID:            | NA   |
| Demographics:        | Country of Origin: United States                             |
| Collection Date:     | 23-DEC-01  |
| Anticoagulant:       | 4% Sodium Citrate  |
| Storage Temperature: | -70 degrees C  |
| Shipping Conditions: | Dry ice  |
| Intended Use:        | For research use only. Not for use in diagnostic procedures. |
| Safety Precautions:  | Use universal precautions when handling this material.       |
| Test Method(s):      | See below  |

| Lot Analysis |         |                          |        |       |
|--------------|---------|--------------------------|--------|-------|
| Test Date    | Analyte | Test Method/Manufacturer | Result | Units |
| 11-Feb-05    | HCV Ag  | Ortho EIA                | 68.775 | S/CO  |

This product was tested for a-HIV and HBsAg and found to be negative.

Data are provided for informational purposes.  
 SeraCare Life Sciences does not claim that others can duplicate test results exactly.  
 NA = Not Available

**Approval:**

|  |                            |
|--|----------------------------|
| <br>_____<br>QA | 05 Jan 15<br>_____<br>Date |
|--|----------------------------|

