

Certificate of Analysis

Product Description: RPR Positive Plasma
 Material No(s): 0365-0005, 0365-0006, 0365-0007, 0365-0008
 Lot No: 9202589
 Donor ID: NA
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: NA
 Anticoagulant: NA
 Storage Temperature: -20
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
10-Jan-07	Rapid Plasma Reagin (Syphilis)	Pulse Agglutination	POS	NA	Neat
10-Jan-07	Rapid Plasma Reagin (Syphilis)	Pulse Agglutination	POS	NA	1:2 dilution
10-Jan-07	Rapid Plasma Reagin (Syphilis)	Pulse Agglutination	POS	NA	1:4 dilution
10-Jan-07	Rapid Plasma Reagin (Syphilis)	Pulse Agglutination	POS	NA	1:8 dilution
10-Jan-07	Rapid Plasma Reagin (Syphilis)	Pulse Agglutination	NEG	NA	1:16 dilution
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIVAg and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
15-Jan-07	Syphilis IgG	Trinity Biotech Captia EIA	2.437	S/CO	NA
04-Jan-08	Syphilis IgG	Trinity Biotech Captia EIA	0.906	S/CO	1:64 dilution
04-Jan-08	Syphilis IgG	Trinity Biotech Captia EIA	0.707	S/CO	1:128 dilution
26-Feb-07	Syphilis IgM	Meddens EIA	2.224	S/CO	NA

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.

NA = Not Available

Approval:

Lynne Rabante

18-Dec-15

Prepared By

Date

Tsering Chelung

22-Dec-15

QA Verified By

Date