

# Certificate of Analysis

Product Description: HCV RNA TYPE 3 Plasma  
 Material No(s): 0310-0057, 0310-0058, 0310-0059  
 Lot No: 9212062  
 Donor ID: NA  
 Gender: NA  
 Age: NA  
 Country of Origin: NA  
 Collection Date: 18-May-06  
 Anticoagulant: NA  
 Storage Temperature: -70°C  
 Shipping Conditions: Dry ice  
 Intended Use: For research use only. Not for use in diagnostic procedures.  
 Safety Precautions: Use universal precautions when handling this material.  
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
29-Oct-07	HCV Genotype	Siemens NS5B Genotyping	3A	NA	NA
01-Nov-15	HCV RNA	Roche COBAS AmpliPrep/TaqMan 2.0	20,535	IU/ML	NA
This product was tested for HBsAg and antibody to HIV-1/2 and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
14-Jun-07	HCV Ab	Ortho EIA	4.732	S/CO	NA

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.  
 NA = Not Available

**Approval:**

*Brenda de Freitas*

28 Dec 15

Prepared By

Date

*Tsering Chelngi*

21 Jan 16

QA Verified By

Date