

Certificate of Analysis

Product Description: HCV RNA TYPE 1 PLASMA
Material No(s): 0310-0039, 0310-0040, 0310-0041
Lot No: 9222515
Donor ID: BD110814
Gender: NA
Age: NA
Country of Origin: Poland
Collection Date: 12-Mar-07
Anticoagulant: NA
Storage Temperature: -70 degrees C
Shipping Conditions: Dry ice
Intended Use: For research use only. Not for use in diagnostic procedures.
Safety Precautions: Use universal precautions when handling this material.
Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
16-Sep-08	HCV Genotype	INNO-LIPA 2.0	1	N/A	
20-Apr-12	HCV RNA	Roche COBAS AmpliPrep/TaqMan	942,000	IU/mL	
20-Apr-12	HCV RNA	Roche COBAS AmpliPrep/TaqMan	1,050,000	IU/mL	

This product was tested for HBsAg and antibody to HIV-1/2 and found to be negative.

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
14-Nov-07	HCV Ab	Ortho EIA	4.967	S/CO	
15-Feb-08	HCV Ab	Abbott EIA	5.917	S/CO	

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Brenda de Freitas

30 Sep 15

Prepared By

Date

[Signature]

30 Sep 15

QA Verified By

Date

