

Certificate of Analysis

Product Description: Chlamydia IgG Positive Plasma
 Material No(s): 0350-0033, 0350-0034, 0350-0035, 0350-0036
 Lot No: 9245832
 Donor ID: BD110121
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: 09-Mar-10
 Anticoagulant: NA
 Storage Temperature: -20
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
12-Oct-12	Chlamydia Pneumonia IgG	DRG International EIA	4.440	S/CO	NA
10-Oct-12	Chlamydia Trachomatis IgG	DRG International EIA	3.649	S/CO	NA
11-May-10	Chlamydia Trachomatis IgG	Trinity Biotech Captia EIA	4.515	S/CO	NA
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIV 1 RNA, HCV RNA and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
10-Oct-12	Chlamydia Pneumonia IgA	DRG International EIA	2.250	S/CO	NA
10-Oct-12	Chlamydia Pneumonia IgM	DRG International EIA	0.344	S/CO	NA
10-Oct-12	Chlamydia Trachomatis IgA	DRG International EIA	1.129	S/CO	NA
10-Oct-12	Chlamydia Trachomatis IgM	DRG International EIA	0.622	S/CO	NA

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Rabante

19-Nov-15

Prepared By

Date

Tsering Chelung

30-Nov-15

QA Verified By

Date