

Certificate of Analysis

Product Description: Lyme IgG Positive Plasma
Material No(s): 0325-0044, 0325-0045, 0325-0046, 0325-0047
Lot No: 9246825
Donor ID: BD110029
Gender: Male
Age: NA
Country of Origin: NA
Collection Date: 13-Jul-10
Anticoagulant: NA
Storage Temperature: -20 degrees C
Shipping Conditions: Dry ice
Intended Use: For research use only. Not for use in diagnostic procedures.
Safety Precautions: Use universal precautions when handling this material.
Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
04-Aug-10	Lyme IgG	Wampole EIA CE Marked	6.757	S/CO	
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIV 1 RNA, HCV RNA, Syphilis and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
10-Aug-10	Lyme IgM	Wampole EIA CE Marked	0.781	S/CO	

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Labonte

13-Oct-15

Prepared By

Date

Christine Pawia

20-Oct-15

QA Verified By

Date

