

Certificate of Analysis

Product Description: Lyme IgM Positive Plasma
 Material No(s): 0325-0048, 0325-0049, 0325-0050, 0325-0051
 Lot No: 9247536
 Donor ID: BD110285
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: 19-Aug-10
 Anticoagulant: NA
 Storage Temperature: -20
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
01-Sep-10	Lyme IgM	Wampole EIA CE Marked	2.612	S/CO	NA
18-Aug-15	Lyme IgM	Zeus EIA	1.400	S/CO	NA
18-Aug-15	Lyme IgM	Zeus EIA	1.656	S/CO	NA
04-Dec-12	Lyme IgM	Wampole EIA CE Marked	1.542	S/CO	1:2 dilution
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIV 1 RNA, HCV RNA, HTLV I/II, and Syphilis and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
18-Aug-15	Lyme IgM	Western Blot	Pos	NA	NA
18-Aug-15	23 Band	Western Blot	Pos	NA	NA
18-Aug-15	39 Band	Western Blot	Neg	NA	NA
18-Aug-15	41 Band	Western Blot	Pos	NA	NA

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:


19-Oct-15

Prepared By

Date



21-Oct-15

QA Verified By

Date

