

Certificate of Analysis

Product Description: Toxoplasma IgM Positive Plasma
 Material No(s): 0320-0007, 0320-0008, 0320-0009, 0320-0010
 Lot No: 9254466
 Donor ID: NA
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: 06-Jun-11
 Anticoagulant: NA
 Storage Temperature: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
18-Oct-11	Toxoplasma IgM	Abbott AxSym	3.072	S/CO	1.843 Index
18-Oct-11	Toxoplasma IgM	Abbott AxSym	1.215	S/CO	0.729 Index (1:4 dilution)
02-May-12	Toxoplasma IgM	Bio-Rad EIA	5.595	S/CO	
02-May-12	Toxoplasma IgM	Bio-Rad EIA	5.298	S/CO	1:4 dilution
09-May-12	Toxoplasma IgM	Bio-Rad EIA	1.458	S/CO	1:8 dilution
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
24-Sep-12	Toxoplasma IgG	Trinity Biotech Captia EIA	7.576	S/CO	NA
05-Oct-12	Toxoplasma IgG	Trinity Biotech Captia EIA	4.299	S/CO	NA

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:


30-Sep-15

Prepared By



Date

02-Oct-15

QA Verified By

Date

