

Certificate of Analysis

Product Description: HTLV-II Antibody Positive Plasma Confirmed
 Material No(s): 0325-0032, 0325-0033, 0325-0034, 0325-0035
 Lot No: 9254598
 Donor ID: BD110991
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: 25-Jun-11
 Anticoagulant: NA
 Storage Temperature: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
18-Oct-11	HTLV-I/II Antibody	Ortho EIA	7.916	S/CO	
02-Nov-11	HTLV-I/II Antibody	Abbott Murex EIA	10.067	S/CO	1:10 dilution
10-Nov-11	HTLV-I/II Antibody	Ortho EIA	8.85	S/CO	1:10 dilution
23-Jan-12	HTLV-I/II Antibody	Avioq EIA	4.395	S/CO	1:10 dilution
23-Jan-12	HTLV-I/II Antibody	Avioq EIA	5.932	S/CO	
27-Dec-12	HTLV-I/II Antibody	MP Diagnostics EIA	8.898	S/CO	1:10 dilution
27-Dec-12	HTLV-I/II Antibody	MP Diagnostics EIA	12.195	S/CO	
24-Oct-11	HTLV-I/II Western Blot	Genelabs Diagnostics Western Blot	HTLV-II	NA	
24-Oct-11	19 Band	Genelabs Diagnostics Western Blot	-	NA	
24-Oct-11	21 Band	Genelabs Diagnostics Western Blot	+	NA	
24-Oct-11	24 Band	Genelabs Diagnostics Western Blot	+	NA	
24-Oct-11	26 Band	Genelabs Diagnostics Western Blot	-	NA	
24-Oct-11	28 Band	Genelabs Diagnostics Western Blot	-	NA	
24-Oct-11	32 Band	Genelabs Diagnostics Western Blot	-	NA	
24-Oct-11	36 Band	Genelabs Diagnostics Western Blot	+	NA	
24-Oct-11	46 Band	Genelabs Diagnostics Western Blot	-	NA	
24-Oct-11	53 Band	Genelabs Diagnostics Western Blot	-	NA	
24-Oct-11	RI Band	Genelabs Diagnostics Western Blot	-	NA	
24-Oct-11	RII Band	Genelabs Diagnostics Western Blot	+	NA	

This product was tested for HBsAg, antibody to HCV, antibody to HIV-1/2, HIV1 RNA, HCV RNA and Syphilis and found to be negative.

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available



Certificate of Analysis

Product Description: HTLV-II Antibody Positive Plasma Confirmed
Material No(s): 0325-0032, 0325-0033, 0325-0034, 0325-0035
Lot No: 9254598

Approval:

Brenda de Freitas

15Sep15

Prepared By

Date

Christine Paiva

22Sep15

QA Verified By

Date