

Certificate of Analysis

Product Description: Toxoplasma IgM Positive Plasma
 Material No(s): 0320-0007, 0320-0008, 0320-0009, 0320-0010
 Lot No: 9255996
 Donor ID: BD110978
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: 14-Dec-11
 Anticoagulant: NA
 Storage Temperature: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
19-Jan-12	Toxoplasma IgM	Abbott AxSym	3.808	S/CO	2.285 Index
19-Jan-12	Toxoplasma IgM	Abbott AxSym	1.393	S/CO	0.836 Index (1:4 dilution)
02-May-12	Toxoplasma IgM	Bio-Rad EIA	1.024	S/CO	NA
02-May-12	Toxoplasma IgM	Bio-Rad EIA	1.202	S/CO	1:4 dilution
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIV 1 RNA, HCV RNA, HBV DNA and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
16-Jan-12	Toxoplasma IgG	Trinity Biotech Captia EIA	5.056	S/CO	NA
16-Jan-12	Toxoplasma IgG	Trinity Biotech Captia EIA	4.99	S/CO	NA

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Labonte

01-Oct-15

Prepared By

Date

[Signature]

02-Oct-15

QA Verified By

Date

