

Certificate of Analysis

Product Description: Toxoplasma IgG Positive Plasma
Material No(s): 0320-0003, 0320-0004, 0320-0005, 0320-0006
Lot No: 9256746
Donor ID: BD110978
Gender: NA
Age: NA
Country of Origin: NA
Collection Date: 12-Feb-12
Anticoagulant: NA
Storage Temperature: -20 degrees C
Shipping Conditions: Dry ice
Intended Use: For research use only. Not for use in diagnostic procedures.
Safety Precautions: Use universal precautions when handling this material.
Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
09-Mar-12	Toxoplasma IgG	Siemens Immulite	110	IU/mL	
24-Sep-12	Toxoplasma IgG	Trinity Biotech Captia EIA	7.576	S/CO	
05-Oct-12	Toxoplasma IgG	Trinity Biotech Captia EIA	4.484	S/CO	
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIV 1 RNA, HCV RNA, HBV DNA, and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
13-Mar-12	Toxoplasma IgM	Bio-Rad EIA	0.927	S/CO	
27-Mar-12	Toxoplasma IgM	Trinity Biotech Captia EIA	0.673	S/CO	

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Laporte

29-Sep-15

Prepared By

[Signature]

Date

02-Oct-15

QA Verified By

Date

