

Certificate of Analysis

Product Description: ToRCH Negative Plasma
 Material No(s): 0320-0011, 0320-0012, 0320-0013, 0320-0014
 Lot No: 9260147
 Donor ID: BD106585
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: 28-May-12
 Anticoagulant: NA
 Storage Temperature: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
20-Aug-12	Toxoplasma IgG	Trinity Biotech Captia EIA	0.085	S/CO	
20-Aug-12	Toxoplasma IgM	Trinity Biotech Captia EIA	0.141	S/CO	
21-Aug-12	Rubella IgG	Wampole EIA	0.016	S/CO	
21-Aug-12	Rubella IgM	Abbott AxSym	0.179	S/CO	
21-Aug-12	CMV IgG	Abbott AxSym	0	S/CO	
21-Aug-12	CMV IgM	Wampole EIA CE Marked	0.208	S/CO	
20-Aug-12	HSV-1/2 IgG	Trinity Biotech Captia EIA	0.035	S/CO	
03-Oct-12	HSV-1/2 IgM	Wampole EIA	0.796	S/CO	
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIV 1 RNA, HCV RNA, Syphilis and found to be negative.					

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Laporte

04-Nov-15

Prepared By

Tsering Chelung

Date

05-Nov-15

QA Verified By

Date