

Certificate of Analysis

Product Description: HTLV-II Antibody Positive Plasma Confirmed
 Material No(s): 0325-0032, 0325-0033, 0325-0034, 0325-0035
 Lot No: 9262240
 Donor ID: NA
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: NA
 Anticoagulant: NA
 Storage Temperature: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
29-Nov-12	HTLV-I/II Antibody	Avioq EIA	1.339	S/CO	1:10 dilution
29-Nov-12	HTLV-I/II Antibody	Avioq EIA	4.336	S/CO	
27-Dec-12	HTLV-I/II Antibody	MP Diagnostics EIA	6.037	S/CO	1:10 dilution
27-Dec-12	HTLV-I/II Antibody	MP Diagnostics EIA	12.195	S/CO	
28-Dec-12	HTLV-I/II LIA	LIA (Ref. Lab)	HTLV-II	NA	
28-Dec-12	19II Band	LIA (Ref. Lab)	POS	NA	
28-Dec-12	46II Band	LIA (Ref. Lab)	POS	NA	
28-Dec-12	gp21 Band	LIA (Ref. Lab)	POS	NA	
28-Dec-12	gp46 Band	LIA (Ref. Lab)	POS	NA	
28-Dec-12	p19 Band	LIA (Ref. Lab)	NEG	NA	
28-Dec-12	P24 Band	LIA (Ref. Lab)	POS	NA	
28-Dec-12	P46I Band	LIA (Ref. Lab)	NEG	NA	
28-Dec-12	STRP Band	LIA (Ref. Lab)	NEG	NA	

This product was tested for HBsAg, antibody to HCV, and antibody to HIV-1/2 and found to be negative.

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:
Brenda de Freitas

15Sep15

Prepared By

Date

Christine Pina

22Sep15

QA Verified By

Date

