

# Certificate of Analysis

Product Description: HTLV-I Antibody Positive Plasma Confirmed  
 Material No(s): 0325-0028, 0325-0029, 0325-0030, 0325-0031  
 Lot No: 9262244  
 Donor ID: NA  
 Gender: NA  
 Age: NA  
 Country of Origin: NA  
 Collection Date: NA  
 Anticoagulant: NA  
 Storage Temperature: -20 degrees C  
 Shipping Conditions: Dry ice  
 Intended Use: For research use only. Not for use in diagnostic procedures.  
 Safety Precautions: Use universal precautions when handling this material.  
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
29-Nov-12	HTLV-I/II Antibody	Avioq EIA	1.345	S/CO	1:10 dilution
29-Nov-12	HTLV-I/II Antibody	Avioq EIA	5.302	S/CO	
27-Dec-12	HTLV-I/II Antibody	MP Diagnostics EIA	6.106	S/CO	1:10 dilution
27-Dec-12	HTLV-I/II Antibody	MP Diagnostics EIA	12.195	S/CO	
28-Dec-12	HTLV-I/II LIA	Reference Lab (Quest. Diag.)	HTLV-I	NA	
28-Dec-12	19II Band	Reference Lab (Quest. Diag.)	Pos	NA	
28-Dec-12	46II Band	Reference Lab (Quest. Diag.)	Neg	NA	
28-Dec-12	gp21 Band	Reference Lab (Quest. Diag.)	Pos	NA	
28-Dec-12	gp46 Band	Reference Lab (Quest. Diag.)	Pos	NA	
28-Dec-12	p19 Band	Reference Lab (Quest. Diag.)	Pos	NA	
28-Dec-12	P24 Band	Reference Lab (Quest. Diag.)	Pos	NA	
28-Dec-12	P46I Band	Reference Lab (Quest. Diag.)	Pos	NA	
28-Dec-12	STRP Band	Reference Lab (Quest. Diag.)	Neg	NA	

This product was tested for HBsAg, antibody to HCV, and antibody to HIV-1/2 and found to be negative.

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.  
 NA = Not Available

**Approval:**

*Brenda de Freitas*

14Sep15

Prepared By

Date

*Christine Paiva*

21Sep15

QA Verified By

Date

