

Certificate of Analysis

Product Description: Chlamydia IgM Positive Plasma
 Material No(s): 0350-0029, 0350-0030, 0350-0031, 0350-0032
 Lot No: BM121029
 Donor ID: NA
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: 23-Sep-00
 Anticoagulant: 4% Sodium Citrate
 Storage Temperature: -20
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis

Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
10-Oct-12	Chlamydia Pneumonia IgM	DRG International EIA	1.004	S/CO	NA
10-Oct-12	Chlamydia Trachomatis IgM	DRG International EIA	2.681	S/CO	NA

This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIVAg, Syphilis and found to be negative.

Additional Data

Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
10-Oct-12	Chlamydia Pneumonia IgA	DRG International EIA	1.714	S/CO	NA
12-Oct-12	Chlamydia Pneumonia IgG	DRG International EIA	0.701	S/CO	NA
10-Oct-12	Chlamydia Trachomatis IgA	DRG International EIA	0.770	S/CO	NA
10-Oct-12	Chlamydia Trachomatis IgG	DRG International EIA	0.361	S/CO	NA

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Labonte

19-Nov-15

Prepared By

Date

Tsering Chelung

01-Dec-15

QA Verified By

Date