

# Certificate of Analysis

Product Description: HSV-1 IgG Positive Plasma  
 Material No(s): 0320-0031, 0320-0032, 0320-0033, 0320-0034  
 Lot No: BM140062  
 Donor ID: BD226093  
 Gender: Female  
 Age: 58  
 Country of Origin: NA  
 Collection Date: 01-Mar-01  
 Anticoagulant: NA  
 Storage Temperature: -20 degrees C  
 Shipping Conditions: Dry ice  
 Intended Use: For research use only. Not for use in diagnostic procedures.  
 Safety Precautions: Use universal precautions when handling this material.  
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
07-Sep-11	HSV-1 IgG	Focus EIA	7.149	S/CO	NA
07-Mar-08	HSV-1 IgG	Trinity Biotech Captia EIA	5.442	S/CO	NA
17-Mar-11	HSV-1 IgG	Trinity Biotech Captia EIA	1.233	S/CO	1:20 dilution
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIVAg and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
08-Apr-10	HSV-1/2 IgG	Siemens Immulite	26.3	S/CO	NA
07-Sep-11	HSV-2 IgG	Focus EIA	0.327	S/CO	NA
12-Mar-08	HSV-2 IgG	Trinity Biotech Captia EIA	1.801	S/CO	NA
12-Mar-08	HSV-1 IgM	Trinity Biotech Captia EIA	1.147	S/CO	NA
12-Mar-08	HSV-2 IgM	Trinity Biotech Captia EIA	1.26	S/CO	NA

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.  
 NA = Not Available

### Approval:

*Lynne Labonte*

06-Oct-15

Prepared By

Date

*Tsering Chelung*

09-Oct-15

QA Verified By

Date

