

Certificate of Analysis

Product Description: HSV-2 IgG Positive Plasma
Material No(s): 0320-0039, 0320-0040, 0320-0041, 0320-0042
Lot No: BM140230
Donor ID: BD226097
Gender: Female
Age: 25
Country of Origin: NA
Collection Date: 05-Sep-02
Anticoagulant: 4% Sodium Citrate
Storage Temperature: -20 degrees C
Shipping Conditions: Dry ice
Intended Use: For research use only. Not for use in diagnostic procedures.
Safety Precautions: Use universal precautions when handling this material.
Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
12-Mar-08	HSV-2 IgG	Trinity Biotech Captia EIA	3.383	S/CO	
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIVAg and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
07-Mar-08	HSV-1 IgG	Trinity Biotech Captia EIA	2.246	S/CO	
12-Mar-08	HSV-1 IgM	Trinity Biotech Captia EIA	0.941	S/CO	
12-Mar-08	HSV-2 IgM	Trinity Biotech Captia EIA	1.427	S/CO	

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Labonte

06-Oct-15

Prepared By

Date

Tsering Chelung

09-Oct-15

QA Verified By

Date

