

Certificate of Analysis

Product Description: Rubella IgG Positive Plasma
 Material No(s): 0320-0023, 0320-0024, 0320-0025, 0320-0026
 Lot No: BM201647
 Donor ID: NA
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: 17-Nov-02
 Anticoagulant: 4% Sodium Citrate
 Storage Temperature: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
20-Sep-05	Rubella IgG	Diamedix EIA Quantitative	4.44	Index	44.4 IU/mL (1:4 dilution)
05-Jan-15	Rubella IgG	Trinity Biotech Captia EIA	2.850	S/CO	1:2 dilution
05-Jan-15	Rubella IgG	Trinity Biotech Captia EIA	2.299	S/CO	1:4 dilution
05-Jan-15	Rubella IgG	Trinity Biotech Captia EIA	1.435	S/CO	1:8 dilution
05-Jan-15	Rubella IgG	Trinity Biotech Captia EIA	1.047	S/CO	1:12 dilution
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HIV 1 RNA, HCV RNA, HIVAg, Syphilis and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
13-Sep-05	Rubella IgM	Diamedix EIA	0.151	S/CO	NA

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Rabonte

05-Oct-15

Prepared By

Date

[Signature]

07-Oct-15

QA Verified By

Date

