

# Certificate of Analysis

Rev. 2

Product Description: HIV-1 Ag Positive Plasma  
 Material No(s): 0315-0033, 0315-0034, 0315-0035  
 Lot No: BM202287  
 Donor ID: N/A  
 Gender: N/A  
 Age: N/A  
 Country of Origin: N/A  
 Collection Date: 10-OCT-02  
 Anticoagulant: N/A  
 Storage Temperature: -20 degrees C  
 Shipping Conditions: Dry ice  
 Intended Use: For research use only. Not for use in diagnostic procedures.  
 Safety Precautions: Use universal precautions when handling this material.  
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
13-APR-10	HIV-1 Antigen	Perkin Elmer EIA	4.652	S/CO	
13-APR-10	HIV-1 Antigen	Zeptometrix EIA	2.677	S/CO	
05-JUL-10	HIV-1 Antigen	Biomerieux VIDAS	14.5	pg/mL	
This product was tested for HBsAg,, and a-HCV and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
20-APR-10	HIV-1/2 Antibody	Ortho Vitros Eci	0.11	S/CO	
20-APR-10	HIV-1/2 Antibody	Ortho Vitros Eci	0.09	S/CO	
15-APR-10	HIV-1/2 Plus O	Abbott PRISM	0.58	S/CO	
15-APR-10	HIV-1/2 Plus O	Abbott PRISM	0.61	S/CO	
07-NOV-05	HIV-1/2 Antibody	Abbott EIA	0.162	S/CO	

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.  
 NA = Not Available

**Approval:**


01-Oct-15

Prepared By

Date



01-Oct-15

QA Verified By

Date

