

Certificate of Analysis

Product Description: CMV IgG Positive Plasma
 Material No(s): 0320-0015, 0320-0016, 0320-0017, 0320-0018
 Lot No: BM204372
 Donor ID: BD101589
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: 03-Dec-05
 Anticoagulant: NA
 Storage Temperature: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
17-Feb-06	CMV IgG	Abbott AxSym	12.12	S/CO	181.8 AU/mL
24-Feb-15	CMV IgG	Trinity Biotech Captia EIA	3.220	S/CO	NA
24-Feb-15	CMV IgG	Trinity Biotech Captia EIA	3.235	S/CO	NA
24-Feb-15	CMV IgG	Trinity Biotech Captia EIA	1.229	S/CO	1:5 dilution
24-Feb-15	CMV IgG	Trinity Biotech Captia EIA	1.251	S/CO	1:5 dilution
24-Feb-15	CMV IgG	Trinity Biotech Captia EIA	0.833	S/CO	1:10 dilution
24-Feb-15	CMV IgG	Trinity Biotech Captia EIA	0.805	S/CO	1:10 dilution
24-Feb-15	CMV IgG	Trinity Biotech Captia EIA	0.520	S/CO	1:20 dilution
24-Feb-15	CMV IgG	Trinity Biotech Captia EIA	0.529	S/CO	1:20 dilution
24-Feb-15	CMV IgG	Trinity Biotech Captia EIA	0.359	S/CO	1:40 dilution
24-Feb-15	CMV IgG	Trinity Biotech Captia EIA	0.350	S/CO	1:40 dilution

This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HTLV I/II, HIV 1 RNA, HCV RNA, HBV DNA and found to be negative.

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
24-Feb-06	CMV IgM	Diamedix EIA	0.764	S/CO	NA

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.

NA = Not Available

Approval:

Lynne Rabonte

01-Oct-15

Prepared By

[Signature]

Date

06-Oct-15

QA Verified By

Date

