

Certificate of Analysis

Product Description: CMV IgG Positive Plasma
Material No(s): 0320-0015, 0320-0016, 0320-0017, 0320-0018
Lot No: BM204443
Donor ID: BD111327
Gender: NA
Age: NA
Country of Origin: NA
Collection Date: 12-Nov-05
Anticoagulant: NA
Storage Temperature: -20 degrees C
Shipping Conditions: Dry ice
Intended Use: For research use only. Not for use in diagnostic procedures.
Safety Precautions: Use universal precautions when handling this material.
Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
03-Mar-06	CMV IgG	Abbott AxSym	15.827	S/CO	237.4 AU/mL
28-Jan-15	CMV IgG	Trinity Biotech Captia EIA	3.066	S/CO	NA
28-Jan-15	CMV IgG	Trinity Biotech Captia EIA	2.225	S/CO	1:2 dilution
28-Jan-15	CMV IgG	Trinity Biotech Captia EIA	1.265	S/CO	1:4 dilution
28-Jan-15	CMV IgG	Trinity Biotech Captia EIA	1.009	S/CO	1:8 dilution
28-Jan-15	CMV IgG	Trinity Biotech Captia EIA	0.575	S/CO	1:12 dilution
28-Jan-15	CMV IgG	Trinity Biotech Captia EIA	0.838	S/CO	1:16 dilution
21-Apr-10	CMV IgG	Siemens Immulite	7.89	Index	NA
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HTLV I/II, Syphilis, HIV 1 RNA, HCV RNA, HBV DNA and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
28-Feb-06	CMV IgM	Trinity Biotech Captia EIA	0.291	S/CO	NA

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Rabonte

01-Oct-15

Prepared By

[Signature]

Date

06-Oct-15

QA Verified By

Date

