

Certificate of Analysis

Product Description: HCV RNA Positive, HCV Ab Negative Plasma
Material No(s): 0310-0036, 0310-0037, 0310-0038
Lot No: BM204839
Donor ID: BD102051
Gender: NA
Age: NA
Country of Origin: NA
Collection Date: 06-Jan-06
Anticoagulant: NA
Storage Temperature: -70 degrees C
Shipping Conditions: Dry ice
Intended Use: For research use only. Not for use in diagnostic procedures.
Safety Precautions: Use universal precautions when handling this material.
Test Method(s): See below

Lot Analysis				
Test Date	Analyte	Test Method/Manufacturer	Result	Units
13-Mar-06	a-HCV	Ortho EIA	0.023	S/CO
17-Aug-14	HCV RNA	Roche COBAS AmpliPrep/Taqman v2.0	56920	IU/mL

This product was tested for HBsAg, a-HIV, and a-HCV and found to be negative.

Data are provided for informational purposes.
 SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:


 QA Prepared By _____ Date 13 Jan 15


 QA Verified By _____ Date 13 Jan 15

