

Certificate of Analysis

Revision 02

| | |
|----------------------|--|
| Product Description: | HIV-1 AB Positive Confirmed Plasma |
| Material No(s): | 0315-0023, 0315-0024, 0315-0025, 0315-0026 |
| Lot No: | BM206402 |
| Donor ID: | BD102915 |
| Demographics: | NA |
| Collection Date: | 25-AUG-06 |
| Anticoagulant: | NA |
| Storage Temperature: | -20 degrees C |
| Shipping Conditions: | Dry ice |
| Intended Use: | For research use only. Not for use in diagnostic procedures. |
| Safety Precautions: | Use universal precautions when handling this material. |
| Test Method(s): | See below |

Lot Analysis

| Test Date | Analyte | Test Method/Manufacturer | Result | Units |
|-----------|--------------------|--------------------------------|--------|-------|
| 16-NOV-06 | Anti-HIV 1/2 | Anti- HIV1/2 Bio-Rad EIA | 9.346 | S/CO |
| 23-JUL-14 | HIV-1 Western Blot | Cambridge Biotech Western Blot | Pos | N/A |
| 23-JUL-14 | BGP120 | Cambridge Biotech Western Blot | 1+ | N/A |
| 23-JUL-14 | BGP160 | Cambridge Biotech Western Blot | 2+ | N/A |
| 23-JUL-14 | BP18 | Cambridge Biotech Western Blot | Neg | N/A |
| 23-JUL-14 | BP24 | Cambridge Biotech Western Blot | Ind | N/A |
| 23-JUL-14 | BP31 | Cambridge Biotech Western Blot | 2+ | N/A |
| 23-JUL-14 | BP41 | Cambridge Biotech Western Blot | 2+ | N/A |
| 23-JUL-14 | BP51 | Cambridge Biotech Western Blot | 2+ | N/A |
| 23-JUL-14 | BP55 | Cambridge Biotech Western Blot | Ind | N/A |
| 23-JUL-14 | BP65/6 | Cambridge Biotech Western Blot | 2+ | N/A |

This product was tested for HBsAg, and HCV and found to be negative.

Data are provided for informational purposes.
SeraCare Life Sciences does not claim that others can duplicate test results exactly.
NA = Not Available

Approval:

Tsering Chelung

QA

10 Dec 15

Date

