

Certificate of Analysis

Product Description: Toxoplasma IgG Positive Plasma
 Material No(s): 0320-0003, 0320-0004, 0320-0005, 0320-0006
 Lot No: BM216493
 Donor ID: NA
 Gender: NA
 Age: NA
 Country of Origin: NA
 Collection Date: 14-Aug-04
 Anticoagulant: NA
 Storage Temperature: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
29-Nov-06	Toxoplasma IgG	Diamedix EIA Quantitative	3.192	Index	159.6 IU/mL
11-Oct-07	Toxoplasma IgG	Abbott AxSym	2.567	S/CO	7.7 IU/mL
11-Oct-07	Toxoplasma IgG	Abbott AxSym	2.767	S/CO	8.3 IU/mL
24-Sep-12	Toxoplasma IgG	Trinity Biotech Captia EIA	6.096	S/CO	NA
05-Oct-12	Toxoplasma IgG	Trinity Biotech Captia EIA	2.457	S/CO	NA
This product was tested for Antibody to HIV 1/2, Antibody to HCV, HBsAg, HTLV I/II and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
29-Dec-06	Toxoplasma IgM	Abbott IMX	0.138	S/CO	NA

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Labonte

29-Sep-15

Prepared By

[Signature]

Date

02-Oct-15

QA Verified By

Date

