

# Certificate of Analysis

Product Description: TPO AB Positive Plasma  
 Material No(s): 0340-0093, 0340-0094, 0340-0095, 0340-0096  
 Lot No: BM216687  
 Donor ID: NA  
 Gender: NA  
 Age: NA  
 Country of Origin: NA  
 Collection Date: NA  
 Anticoagulant: NA  
 Storage Temperature: -20 degrees C  
 Shipping Conditions: Dry ice  
 Intended Use: For research use only. Not for use in diagnostic procedures.  
 Safety Precautions: Use universal precautions when handling this material.  
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
30-Mar-10	Thyroid Peroxidase Antibody	Siemens Immulite	55.5	IU/ML	1:10 dilution
19-Jan-07	Thyroid Peroxidase Antibody	Diamedix EIA	525.7	IU/ML	

This product was tested for HBsAg, Antibody to HIV 1/2, Antibody to HCV, HIVAg, and Antibody to HTLV I/II and found to be negative.

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
21-Feb-07	Thyroglobulin Antibody	Diamedix EIA	12.3	IU/ML	
31-Mar-10	Thyroglobulin Antibody	Siemens Immulite	2.58	IU/ML	
04-Feb-13	C Reactive Protein	Beckman Immage	0.267	MG/DL	

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.  
 NA = Not Available

**Approval:**

*Lynne Rabente*

21-Jul-15

Prepared By

Date

*Christine Pawia*

29-Jul-15

QA Verified By

Date

