

Syphilis Mixed Titer AccuSet™ Performance Panel Modified PSS202(M2)

A SERACARE PANEL PRODUCT

INTENDED USE

The Syphilis Mixed Titer AccuSet™ Performance Panel Modified is intended for use by diagnostic manufacturers, researchers, and clinical laboratories to develop, evaluate, or troubleshoot syphilis test methods. Characterized samples and comprehensive data are provided for comparative analysis. For research use only. Not for use in diagnostic procedures.

PRODUCT DESCRIPTION

This product is a 17-member panel originating from Syphilis Mixed Titer Performance Panel PSS202-0.5. Panel members 1, 4 and 16 from the original 20-member panel are no longer available. Panel members are undiluted, naturally occurring plasma samples with reactivity ranging from negative to highly positive for treponema antibody and reagin. Units were aseptically filtered. No preservative was added.

Cat. No. PSS202 (M2)-0.5 / Cross Ref. No. 0820-0191
1 vial per member
17 members, 0.5 mL per vial

STORAGE

Panel members should be stored at -20°C or colder. SeraCare Life Sciences recommends that the panel members be divided into smaller aliquots to avoid multiple freeze-thaw cycles, if appropriate. If turbidity or particulate matter is observed, the samples should be centrifuged in accordance with each test kit manufacturer's instructions for sample preparation.

INTERPRETATION OF RESULTS

The Data Sheet for Syphilis Mixed Titer AccuSet™ Performance Panel PSS202(M2) is available at www.seracare.com. The Data Sheet lists results for panel members generated using commercially-available

screening, monitoring, and confirmatory test methods. Tests were performed at SeraCare or at recognized reference laboratories (RL) by individuals who routinely use these procedures. Information regarding specific test methods is available on the Data Sheet. Data Sheets are updated when new data are available.

LIMITATIONS

The Syphilis Mixed Titer AccuSet™ Performance Panel Modified is offered for research use only, not for use in diagnostic procedures. Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.

PRECAUTIONS

Follow Universal Precautions.¹ The units that make up this panel were tested and found negative for HBsAg, anti-HIV, and anti-HCV. This does not ensure the absence of these or other human pathogens.

Do not pipette by mouth. Do not smoke, eat, or drink in areas where specimens are handled.

These materials should be disposed of in a manner that will inactivate pathogenic agents.

REFERENCES

1. Siegel JD, Rhinehart E, Jackson M, Chiarello L, and the Healthcare Infection Control Practices Advisory Committee, 2007 Guideline for Isolation Precautions: Preventing Transmission of Infectious Agents in Healthcare Settings

For assistance, contact SeraCare Technical Support at 508.244.6400.

The Data Sheet for this panel in PDF form can be found at www.seracare.com

A printed copy of the Data Sheet may be requested by email at info@seracare.com, or by phone at 508.244.6400

ASK ABOUT RELATED SERACARE PRODUCTS

- ACCURUN® Quality Controls
- AccuType™ Viral Isolates
- SeraCon™ and Basematrix Processed Plasma

