



# Refractrol-SP Serum Protein Assay Sheet

Product Description: Refractrol-SP Low  
Material Number: 84652  
Batch: 10735156  
Expiration Date: 21 MAR 2028  
Storage Temperature: Ambient Room Temperature (typically 15-30 °C)

Assay values obtained for Reichert Digital and Analog models temperature compensated:

Average Concentration	Uncertainty	Units
4.8	+/- 0.4	g/dL


QC Testing:


Assay values obtained for Atago models when operated at room temperature:

Average Concentration	Uncertainty	Units
4.5	+/- 0.6	g/dL

Safety Precautions: Use Universal Precautions when handling this material.

Approval:

  
Prepared By \_\_\_\_\_ Date 19 JUN 2025

  
QA Verified By \_\_\_\_\_ Date 19 JUN 2025