

Certificate of Analysis

Revised 28 OCT 2022

Product Description: HBs AB Positive Greater Than 1000 mIU/mL Plasma
Material Number: 0305-0049, 0305-0050, 0305-0051, 0305-0052
Batch: 10067693
Donor ID: BD216262
Gender: NA
Age: NA
Country of Origin: United States
Collection Date: 15 MAY 2012
Anticoagulant: 4% Sodium Citrate
Storage: -20 degrees C
Shipping Conditions: Dry ice
Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.
Safety Precautions: Use universal precautions when handling this material.
Test Method(s): See below

Data are provided for informational purposes.
LGC Clinical Diagnostics does not claim that others can duplicate the test results indicated below.

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
20 Jan 2015	HBs Antibody (anti-HBs)	Abbott Architect	1,000.000	mIU/mL	ALR
07 Aug 2017	HBs Antibody (anti-HBs)	DiaSorin EIA Plus	20.619	S/CO	Neat
07 Aug 2017	HBs Antibody (anti-HBs)	DiaSorin EIA Plus	20.619	S/CO	1:25 dilution
07 Aug 2017	HBs Antibody (anti-HBs)	DiaSorin EIA Plus	20.619	S/CO	1:50 dilution
07 Aug 2017	HBs Antibody (anti-HBs)	DiaSorin EIA Plus	20.619	S/CO	1:100 dilution
03 Nov 2014	HBs Antibody (anti-HBs)	Ortho Vitros Eci	>1,000.000	mIU/mL	NA

This product was tested for Antibodies to HIV 1/2, Antibodies to HCV, Antibodies to HTLV I/II, HBsAg and found to be negative.

NA = Not Available


Approval:



28 OCT 2022

Prepared By

Date



28 OCT 2022

QA Verified By

Date