

Certificate of Analysis

Product Description: Material Number: Batch:	HBsAg Positive Plasma (Quantitative) 0305-0061, 0305-0062, 0305-0063, 0305-0064 9103149
Donor ID:	NA
Gender:	NA
Age:	NA
Country of Origin:	Brazil
Collection Date:	NA
Anticoagulant:	CPDA
Storage:	-20 degrees C
Shipping Conditions:	Dry ice
Intended Use:	For research use or for further manufacture into in-vitro diagnostic reagents.
Safety Precautions: Test Method(s):	Use universal precautions when handling this material. See below

Data are provided for informational purposes.

LGC Clinical Diagnostics does not claim that others can duplicate the test results indicated below.

Batch Analysis							
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes		
18 Aug 2022	HBsAg	Abbott Architect	1,821.390	S/CO	NA		
					≥0.05 IU/MLis		
27 Jan 2023	HBsAg	Abbott Architect Europe	52.480	IU/ML	Pos		
This product was tested for Antibodies to HIV 1/2, Antibodies to HCV and found to be negative.							

Additional Data								
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes			
19 Aug 2022	Hepatitis Delta IgG	DiaSorin Liaison XL	<0.100	AU/ML	NA			

NA = Not Available

Approval:

Lynne Laborte

Prepared By

Jaula J. Kuman

QA Verified By

03 FEB 2023 Date

08 FEB 2023

Date