

Certificate of Analysis

Product Description: Material Number: Batch:	HBsAg Positive Plasma (Quantitative) 0305-0061, 0305-0062, 0305-0063, 0305-0064 9118603
Donor ID:	BD110425
Gender:	NA
Age:	NA
Country of Origin:	NA
Collection Date:	22 JUL 2000
Anticoagulant:	4% Sodium Citrate
Storage:	-20 degrees C
Shipping Conditions:	Dry ice
Intended Use:	For research use or for further manufacture into in-vitro diagnostic reagents.
Safety Precautions: Test Method(s):	Use universal precautions when handling this material. See below

Data are provided for informational purposes.

LGC Clinical Diagnostics does not claim that others can duplicate the test results indicated below.

Batch Analysis							
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes		
10 Feb 2022	HBsAg	Abbott Architect	10.110	S/CO	NA		
					≥0.05 IU/MLis		
27 Jan 2023	HBsAg	Abbott Architect Europe	0.320	IU/ML	Pos		
This product was tested for Antibodies to HIV 1/2, Antibodies to HCV and found to be negative.							

Additional Data								
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes			
10 Feb 2022	HBc IgM	Abbott Architect	0.160	S/CO	NA			

NA = Not Available

Approval:

Lynne Laboute

Prepared By

Jaula J. Kuman

QA Verified By

03 FEB 2023 Date

08 FEB 2023

Date