

Certificate of Analysis

Rev. 02

Product Description: HCV Ab Positive Confirmed Plasma
Material No(s): 0310-0023, 0310-0024, 0310-0025, 0310-0026
Lot No: 9249846
Donor ID: BD110405
Gender: Male
Age: 50
Country of Origin: NA
Collection Date: 04-Oct-10
Anticoagulant: NA
Storage Temperature: -20 degrees C
Shipping Conditions: Dry ice
Intended Use: For research use only. Not for use in diagnostic procedures.
Safety Precautions: Use universal precautions when handling this material.
Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
30-Jan-11	RIBA band c100	Chiron RIBA HCV 3.0 SIA	4+	NA	
30-Jan-11	RIBA band c22	Chiron RIBA HCV 3.0 SIA	4+	NA	
30-Jan-11	RIBA band c33	Chiron RIBA HCV 3.0 SIA	4+	NA	
30-Jan-11	RIBA band NS5	Chiron RIBA HCV 3.0 SIA	3+	NA	
30-Jan-11	RIBA band SOD	Chiron RIBA HCV 3.0 SIA	-	NA	
18-Jan-11	a-HCV	Ortho EIA	4.926	S/CO	
30-Jan-11	HCV RIBA	Chiron RIBA HCV 3.0 SIA	Positive	NA	
This product was tested for HBsAg, a-HIV 1/2 and found to be negative.					

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.

NA = Not Available

Approval:

06-Oct-15

Prepared By

Date



06-Oct-15

QA Verified By

Date

