

Certificate of Analysis

Revised 06APR2018

Product Description: HIV 1/2 AB Positive Plasma
 Material No(s): 0315-0019, 0315-0020, 0315-0021, 0315-0022
 Lot No: 9259400
 Donor ID: NA
 Gender: Male
 Age: 23
 Country of Origin: Ghana
 Collection Date: 22-Mar-12
 Anticoagulant: CPDA-1
 Storage Temperature: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
08-Jun-12	HIV-1/2 Antibody	Bio-Rad EIA	10.417	S/CO	NA
12-Jun-12	HIV-1 p24 Ag	Perkin Elmer EIA	0.258	S/CO	NA
This product was tested for Antibody to HCV, HBsAg and found to be negative.					

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available


Approval:



06 APR 2018

Prepared By

Date



18 APR 2018

QA Verified By

Date