

Certificate of Analysis

Product Description: HIV-2 AB Positive Confirmed Plasma
 Material Number: 0315-0027, 0315-0028, 0315-0029
 Batch: 10266760
 Donor ID: BD200052
 Gender: NA
 Age: NA
 Country of Origin: Germany
 Collection Date: 19 MAR 2014
 Anticoagulant: EDTA
 Storage: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Data are provided for informational purposes.

SeraCare Life Sciences does not claim that others can duplicate the test results indicated below.

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
28 Oct 2017	gp140 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28 Oct 2017	GP160 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28 Oct 2017	gp36 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28 Oct 2017	gp41 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28 Oct 2017	P24 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	NEG	NA	NA
28 Oct 2017	P31 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28 Oct 2017	HIV-1 Antibody	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28 Oct 2017	HIV-2 Antibody	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	HIV2 Pos with HIV1 Cross Reactivity
28 Nov 2017	gp140 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28 Nov 2017	GP160 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28 Nov 2017	gp36 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28 Nov 2017	gp41 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28 Nov 2017	P24 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	NEG	NA	NA
28 Nov 2017	P31 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28 Nov 2017	HIV-1 Antibody	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	HIV2 Pos with HIV1 Cross Reactivity

Certificate of Analysis

Product Description: HIV-2 AB Positive Confirmed Plasma
 Material Number: 0315-0027, 0315-0028, 0315-0029
 Batch: 10266760

28 Nov 2017	HIV-2 Antibody	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
23 Jun 2017	HIV-2 Antibody	Bio-Rad EIA	18.857	S/CO	Neat
23 Jun 2017	HIV-2 Antibody	Bio-Rad EIA	11.057	S/CO	1:50 dilution
This product was tested for Antibodies to HCV, Antibodies to HTLV I/II, HBsAg and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
26 Jun 2017	HIV-1 Antibody	Avioq EIA	2.291	S/CO	NA
20 Oct 2017	HIV-1/2 Plus O Antibody	Bio-Rad EIA	14.440	S/CO	OFF SCALE

NA = Not Available

Approval:

Lynne Labonte

21 OCT 2021

Prepared By

Date

Danni Parker

25 OCT 2021

QA Verified By

Date