

# Certificate of Analysis

Product Description: HIV-2 AB Positive Confirmed Plasma  
 Material Number: 0315-0027, 0315-0028, 0315-0029  
 Batch: 10266767  
 Donor ID: BD200052  
 Gender: NA  
 Age: NA  
 Country of Origin: Germany  
 Collection Date: 18 NOV 2013  
 Anticoagulant: NA  
 Storage: -20 degrees C  
 Shipping Conditions: Dry ice  
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.  
 Safety Precautions: Use universal precautions when handling this material.  
 Test Method(s): See below

**Data are provided for informational purposes.**

**SeraCare Life Sciences does not claim that others can duplicate the test results indicated below.**

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
23-Jun-17	HIV-2 Antibody	Bio-Rad EIA	9.966	S/CO	1:50 dilution
23-Jun-17	HIV-2 Antibody	Bio-Rad EIA	18.394	S/CO	Neat
28-Oct-17	gp140 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28-Oct-17	GP160 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28-Oct-17	gp36 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28-Oct-17	gp41 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28-Oct-17	P24 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	NEG	NA	NA
28-Oct-17	P31 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28-Oct-17	HIV-2 Antibody	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	HIV2 Pos with HIV1 Cross Reactivity
28-Oct-17	HIV-1 Antibody	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28-Nov-17	gp140 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28-Nov-17	GP160 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28-Nov-17	gp36 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	NEG	NA	NA
28-Nov-17	gp41 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28-Nov-17	P24 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	NEG	NA	NA

# Certificate of Analysis

Product Description: HIV-2 AB Positive Confirmed Plasma  
 Material Number: 0315-0027, 0315-0028, 0315-0029  
 Batch: 10266767

28-Nov-17	P31 Band	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
28-Nov-17	HIV-2 Antibody	Bio-Rad Geenius HIV1/2 Confirmatory Assay	IND	NA	NA
28-Nov-17	HIV-1 Antibody	Bio-Rad Geenius HIV1/2 Confirmatory Assay	POS	NA	NA
This product was tested for Antibodies to HCV, Antibodies to HTLV I/II, HBsAg and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
26-Jun-17	HIV-1 Antibody	Avioq EIA	2.302	S/CO	NA
20-Oct-17	HIV-1/2 Plus O Antibody	Bio-Rad EIA	14.440	S/CO	NA

NA = Not Available

**Approval:**

*Lynne Labonte*

21 JUL 2020

Prepared By

Date

*Dianne Parker*

24 JUL 2020

QA Verified By

Date