

# Certificate of Analysis

Product Description: HIV-2 AB Positive Confirmed Plasma  
 Material Number: 0315-0027, 0315-0028, 0315-0029  
 Batch: 10682126  
 Donor ID: BD519646  
 Gender: Male  
 Age: N/A  
 Country of Origin: United States  
 Collection Date: 12 APR 2023  
 Anticoagulant: 4% Sodium Citrate  
 Storage: -20 degrees C  
 Shipping Conditions: Dry ice  
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.  
 Safety Precautions: Use universal precautions when handling this material.  
 Test Method(s): See below


**Data are provided for informational purposes. LGC Clinical Diagnostics does not claim that others can duplicate the test results indicated below.**


Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
21 Nov 2023	gp140 Band	Bio-Rad Geenius	Positive	NA	NA
21 Nov 2023	GP160 Band	Bio-Rad Geenius	Negative	NA	NA
21 Nov 2023	gp36 Band	Bio-Rad Geenius	Positive	NA	NA
21 Nov 2023	gp41 Band	Bio-Rad Geenius	Positive	NA	NA
21 Nov 2023	P24 Band	Bio-Rad Geenius	Negative	NA	NA
21 Nov 2023	P31 Band	Bio-Rad Geenius	Positive	NA	NA
21 Nov 2023	HIV-1 Antibody	Bio-Rad Geenius	Positive.	NA	HIV-1 Cross reactivity
21 Nov 2023	HIV-2 Antibody	Bio-Rad Geenius	Positive	NA	NA
14 Jun 2024	HIV-2 Antibody	Bio-Rad EIA	12.678	S/CO	Neat
14 Jun 2024	HIV-2 Antibody	Bio-Rad EIA	2.948	S/CO	1:50 dilution
23 Apr 2024	HIV-2 Antibody	BioPlex 2200	RESULT >200.00	INDEX	N/A
23 Apr 2024	HIV-1 Antibody	BioPlex 2200	0.690	INDEX	NA
23 Apr 2024	HIV-1 p24 Ag	BioPlex 2200	0.200	INDEX	NA
23 Apr 2024	HIV-1/2 Ag/Ab	BioPlex 2200	RESULT >200.00	INDEX	N/A
28 Aug 2023	HIV-1 Antibody	Aviog EIA	3.127	S/CO	NA
25 Aug 2023	HIV-1/2 Ag/Ab	Abbott Architect	174.420	S/CO	N/A

This product was tested for HBsAg, Antibodies to HCV, Antibodies to HTLV I/II, and found to be negative.

NA = Not Available

**Approval:**


22 Apr 2026  
 Prepared By Date


27 APR 2026  
 QA Verified By Date