

Certificate of Analysis

Rev 2

Product Description: HIV-2 AB Positive, Confirmed, HIV-1 AB Negative Plasma
 Material No(s): 0315-0030, 0315-0031, 0315-0032
 Lot No: 10101738
 Donor ID: BD200481
 Gender: Male
 Age: NA
 Country of Origin: US
 Collection Date: 21-Jan-15
 Anticoagulant: 4% Sodium Citrate
 Storage Temperature: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use only. Not for use in diagnostic procedures.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Lot Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
23-Feb-15	HIV-2 Rapid Test	Bio-Rad Rapid Test (Multispot)	POS	NA	NA
24-Feb-15	HIV-2 Antibody	Bio-Rad EIA	5.602	S/CO	NA
This product was tested for Antibody to HCV, HBsAg and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
23-Feb-15	HIV-1 Antibody	Avioq EIA	0.677	S/CO	NA
23-Feb-15	HIV-1 Rapid Test	Bio-Rad Rapid Test (Multispot)	NEG	NA	NA
20-Sep-16	HIV-1/2 Plus O Antibody	Bio-Rad EIA	14.870	S/CO	OFF SCALE

Data are provided for informational purposes. SeraCare Life Sciences does not claim that others can duplicate test results exactly.
 NA = Not Available

Approval:

Lynne Rabente

03-Jan-17

Prepared By

Date

Buwa

05-Jan-17

QA Verified By

Date