

Certificate of Analysis

Product Description: HIV-2 AB Positive, Confirmed HIV-1 AB Negative Plasma
 Material Number: 0315-0030, 0315-0031, 0315-0032
 Batch: 10624547
 Donor ID: BD406473
 Gender: NA
 Age: NA
 Country of Origin: Germany
 Collection Date: 05 AUG 2012
 Anticoagulant: 4% Sodium Citrate
 Storage: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Data are provided for informational purposes.

LGC Clinical Diagnostics does not claim that others can duplicate the test results indicated below.

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
05 Mar 2024	gp140 Band	Bio-Rad Geenius	Positive	NA	NA
05 Mar 2024	GP160 Band	Bio-Rad Geenius	Negative	NA	NA
05 Mar 2024	gp36 Band	Bio-Rad Geenius	Positive	NA	NA
05 Mar 2024	gp41 Band	Bio-Rad Geenius	Negative	NA	NA
05 Mar 2024	P24 Band	Bio-Rad Geenius	Negative	NA	NA
05 Mar 2024	P31 Band	Bio-Rad Geenius	Positive	NA	NA
05 Mar 2024	HIV-1 Antibody	Bio-Rad Geenius	Indeterminate	NA	NA
05 Mar 2024	HIV-2 Antibody	Bio-Rad Geenius	Positive	NA	NA
16 Jun 2022	HIV-2 Antibody	Bio-Rad EIA	15.324	S/CO	Neat
16 Jun 2022	HIV-2 Antibody	Bio-Rad EIA	7.728	S/CO	1:50 dilution
14 Jun 2022	HIV-1 Antibody	Avioq EIA	0.548	S/CO	NA
This product was tested for Antibodies to HCV, Antibodies to HTLV I/II, HBsAg and found to be negative.					

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
14 Jun 2022	HIV-1/2 Ag/Ab	Abbott Architect	173.040	S/CO	NA

NA = Not Available

Approval:

Lynne Rabente

02 OCT 2024

Prepared By

Date

Gaula J. Kuman

04 OCT 2024

QA Verified By

Date