

# Certificate of Analysis

**Revised 13JAN2022**

Product Description: Toxoplasma IgM Positive Plasma  
 Material Number: 0320-0007, 0320-0008, 0320-0009, 0320-0010  
 Batch: 9254466  
 Donor ID: NA  
 Gender: NA  
 Age: NA  
 Country of Origin: NA  
 Collection Date: 06 JUN 2011  
 Anticoagulant: NA  
 Storage: -20 degrees C  
 Shipping Conditions: Dry ice  
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.  
 Safety Precautions: Use universal precautions when handling this material.  
 Test Method(s): See below

**Data are provided for informational purposes.  
 SeraCare Life Sciences does not claim that others can duplicate the test results indicated below.**

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
02 May 2012	Toxoplasma IgM	Bio-Rad EIA	5.595	S/CO	Neat
02 May 2012	Toxoplasma IgM	Bio-Rad EIA	5.298	S/CO	1:4 dilution
09 May 2012	Toxoplasma IgM	Bio-Rad EIA	1.458	S/CO	1:8 dilution
25 Feb 2020	Toxoplasma IgM	DiaSorin Liaison	34.800	AU/ML*	Neat
25 Feb 2020	Toxoplasma IgM	DiaSorin Liaison	8.060	AU/ML*	1:4 dilution

This product was tested for Antibodies to HIV 1/2, Antibodies to HCV, HBsAg and found to be negative.  
 \*≥8 AU/ML is positive based on site utilized

Additional Data					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
05 Oct 2012	Toxoplasma IgG	Trinity Biotech Captia EIA	4.299	S/CO	NA
01 Oct 2018	Toxoplasma IgG	Trinity Biotech Captia EIA	4.990	S/CO	Neat
01 Oct 2018	Toxoplasma IgG	Trinity Biotech Captia EIA	3.125	S/CO	1:10 dilution

NA = Not Available

**Approval:**

*Lynne Labonte*

13 JAN 2022

Prepared By

Date

*Danni Parker*

20 JAN 2022

QA Verified By

Date