

Certificate of Analysis

Product Description: HTLV I/II Antibody Positive Plasma
 Material Number: 0325-0036, 0325-0037, 0325-0038, 0325-0039
 Batch: 10612550
 Donor ID: BD494821
 Gender: Male
 Age: 33
 Country of Origin: Honduras
 Collection Date: 02 OCT 2021
 Anticoagulant: CPDA-1
 Storage: -20 degrees C
 Shipping Conditions: Dry ice
 Intended Use: For research use or for further manufacture into in-vitro diagnostic reagents.
 Safety Precautions: Use universal precautions when handling this material.
 Test Method(s): See below

Data are provided for informational purposes.
LGC Clinical Diagnostics does not claim that others can duplicate the test results indicated below.

Batch Analysis					
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes
27 Apr 2022	gp21 Band	INNO-LIA	POS	NA	NA
27 Apr 2022	gp46 Band	INNO-LIA	NEG	NA	NA
27 Apr 2022	46I Band	INNO-LIA	NEG	NA	NA
27 Apr 2022	46II Band	INNO-LIA	NEG	NA	NA
27 Apr 2022	p19 Band	INNO-LIA	NEG	NA	NA
27 Apr 2022	p 19 I/II Band	INNO-LIA	POS	NA	NA
27 Apr 2022	P24 Band	INNO-LIA	POS	NA	NA
27 Apr 2022	HTLV-I/II Antibody	INNO-LIA	POS	NA	UNTYPEABLE
08 Apr 2022	HTLV-I/II Antibody	Avioq EIA	8.564	S/CO	Neat
08 Apr 2022	HTLV-I/II Antibody	Avioq EIA	8.283	S/CO	1:10 dilution
This product was tested for Antibodies to HIV 1/2, Antibodies to HCV, HBsAg and found to be negative.					

NA = Not Available

Approval:

Sydney Labonte

05 MAY 2022

Prepared By

Date

Danni Parker

09 MAY 2022

QA Verified By

Date