

Certificate of Analysis

Revised 22 NOV 2022	-
Product Descripti	on: Syphilis (T. Pallidum) IgG Positive Plasma
Material Number:	0365-0001, 0365-0002, 0365-0003, 0365-0004
Batch:	10267131
Donor ID:	BD328562
Gender:	NA
Age:	NA
Country of Origin	United States
Collection Date:	17 MAY 2016
Anticoagulant:	4% Sodium Citrate
Storage:	-20 degrees C
Shipping Condition	ons: Dry ice
Intended Use:	For research use or for further manufacture into in-vitro diagnostic reagents.
Safety Precautior	ns: Use universal precautions when handling this material.
Test Method(s):	See below

Data are provided for informational purposes.

LGC Clinical Diagnostics does not claim that others can duplicate the test results indicated below.

Batch Analysis							
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes		
03 Oct 2019	Syphilis Total Ab	Abbott Architect	18.070	S/CO	NA		
14 Aug 2017	Syphilis IgG	Trinity Biotech Captia EIA	1.596	S/CO	NA		
This product was tested for Antibodies to HIV 1/2, Antibodies to HCV, HBsAg, HIV 1 RNA, HCV RNA, HBV DNA and found to be negative.							

Additional Data						
Test Date	Analyte	Test Method/Manufacturer	Result	Units	Notes	
09 Oct 2019	Rapid Plasma Reagin (Syphilis)	Becton Dickinson RPR	POS	NA	Neat	
	Rapid Plasma Reagin					
09 Oct 2019	(Syphilis)	Becton Dickinson RPR	POS	NA	1:2 dilution	
09 Oct 2019	Rapid Plasma Reagin (Syphilis)	Becton Dickinson RPR	POS	NA	1:4 dilution	
09 Oct 2019	Rapid Plasma Reagin (Syphilis)	Becton Dickinson RPR	NEG	NA	1:8 dilution	

NA = Not Available

Approval:

Lynne Laborte

11 NOV 2022

Prepared By

Jaula J. Kuman

QA Verified By

22 NOV 2022

Date

Date

www.seracare.com