

# Certificate of Analysis

## Research Use Only - Not for Use in Diagnostic Procedures

Product Description:	<b>Product</b>	<b>Material Code</b>	<b>Batch Number</b>
	AccuPlex™ SARS-CoV-2 Verification Panel	0505-0129	10482502

Date of Manufacture: 06 APR 2020

Expiration Date: 06 APR 2022

Storage Temperature: 2 - 8 °C



Kit Component	Qualitative Result
SARS-CoV-2 Member 1	Positive
SARS-CoV-2 Member 2	Positive
SARS-CoV-2 Member 3	Positive
SARS-CoV-2 Negative	Negative

QC Testing:

Real time qPCR testing using nCoV 2019 primers/probes and protocol described in US Centers for Disease Control Assay publication and instructions for use<sup>1</sup>.

Safety Precautions: Preparation contains recombinant Sindbis virus. Use the Centers for Disease Control recommended universal precautions.

### Approval:

	08 APR 2020
Prepared By	Date
	08 APR 2020
QA Verified By	Date

<sup>1</sup> CDC 2019-Novel Coronavirus (2019-nCoV) Real Time RT-PCR Diagnostic Panel Instructions for use. CDC-006-00019 Revision: 01. Effective 2/4/2020.