

# Revised Certificate of Analysis

## Research Use Only - Not for Use in Diagnostic Procedures

Product Description:

Product	Material Number	Batch Number
AccuPlex™ SARS-CoV-2 Verification Panel - Full Genome	0505-0168	10597816

Date of Manufacture: 17 DEC 2021  
 Expiration Date: 10 MAR 2027  
 Storage Temperature: 2 - 8 °C

Real time qPCR testing using nCoV 2019 primers/probes and protocol described in US Centers for Disease Control Assay publication and instructions for use<sup>1</sup>. Average cycle threshold values for Panel Members 1-3 are consistent with 10-fold serial dilution.

QC Testing:

Kit Component	Qualitative Result
SARS-CoV-2 Member 1	Positive
SARS-CoV-2 Member 2	Positive
SARS-CoV-2 Member 3	Positive
SARS-CoV-2 Negative	Negative

Safety Precautions:

Preparation contains recombinant Sindbis virus. Use the Centers for Disease Control recommended universal precautions.

<sup>1</sup>CDC 2019-Novel Coronavirus (2019-nCoV) Real Time RT-PCR Diagnostic Panel Instructions for use. CDC-006-00019 Revision: 03. Effective 330/2020.

Approval:

Prepared By  Date 18-MAR-2025

QA Verified By  Date 19 Mar 2025