

Certificate of Analysis

For In Vitro Diagnostic Use

| Product Description: | AccuPlex™ SARS-CoV-2 Full Genome w/cRNA |
|----------------------|---|
| | Positive Control 1500 on/ml Kit |

Positive Control 1500 cp/ml Kit

Material Number: 0505-0220

Kit/Vial Batch Number: 10576953

10x Bulk Batch Number: 10583164

Concentration of SARS-CoV-2 N region

at 10X bulk stage (as determined by

digital PCR)

15,560 copies/mL

Qualitative Result: Real time qPCR testing using nCoV 2019 primers/probes and protocol described in US Centers for

Disease Control Assay publication and

instructions for use1.

Safety Precautions:

Positive

Nominal Carrier RNA Concentration: 5 μg/mL

Expiration Date: 2023-10-16

Storage Temperature: 2 - 8 °C

Volume per vial: 0.3 mL

Preparation contains recombinant Sindbis virus. Use the

US Centers for Disease Control recommended universal precautions for handling the material.

Approval:

Koj 9 5to A 24 NOV 2021

Prepared By Date

24 NOV 2021

QA Verified By Date

CDC 2019-Novel Coronavirus (2019-nCoV) Real Time RT-PCR Diagnostic Panel Instructions for use. CDC-006-00019 Revision: 06. Effective 12/01/2020.