



Certificate of Analysis

For In Vitro Diagnostic Use

| | |
|--|---|
| Product Description: | AccuPlex™ SARS-CoV-2 Full Genome w/cRNA Positive Control 1500 cp/ml Kit |
| Material Number: | 0505-0220 |
| Kit/Vial Batch Number: | 10576956 |
| 10x Bulk Batch Number: | 10583164 |
| Concentration of SARS-CoV-2 N region at 10X bulk stage (as determined by digital PCR) | 15,560 copies/mL |
| Qualitative Result: Real time qPCR testing using nCoV 2019 primers/probes and protocol described in US Centers for Disease Control Assay publication and instructions for use ¹ . | Positive |
| Nominal Carrier RNA Concentration: | 5 µg/mL |
| Expiration Date: | 2023-10-16 |
| Storage Temperature: | 2 - 8 °C |
| Volume per vial: | 0.3 mL |
| Safety Precautions: | Preparation contains recombinant Sindbis virus. Use the US Centers for Disease Control recommended universal precautions for handling the material. |

Approval:

Prepared By

09 DEC 2021

Date

QA Verified By

09 DEC 2021

Date

¹ CDC 2019-Novel Coronavirus (2019-nCoV) Real Time RT-PCR Diagnostic Panel Instructions for use. CDC-006-00019 Revision: 06. Effective 12/01/2020.