



Certificate of Analysis

For In Vitro Diagnostic Use

Product Description:	AccuPlex™ Human RNase P Negative Control Kit
Material Number:	0505-0242
Kit Batch Number:	10551261
Bulk Batch Number:	10533704
Concentration of RNase P at bulk stage (as determined by digital PCR)	5600 copies / mL
Qualitative Result: Real time qPCR testing using nCoV 2019 primers/probes and protocol described in US Centers for Disease Control Assay publication and instructions for use ¹ .	Negative
Expiration Date:	2022-12-11
Storage Temperature:	2-8 °C
Volume per Vial:	0.4 mL
Safety Precautions:	Preparation contains recombinant Sindbis virus. Use the US Centers for Disease Control recommended universal precautions for handling the material.

Approval:

28 APR 2021

Prepared By

Date

29 APR 2021

QA Verified By

Date

¹ CDC 2019-Novel Coronavirus (2019-nCoV) Real Time RT-PCR Diagnostic Panel Instructions for use.
CDC-006-00019 Revision: 06. Effective 12/01/2020.