

Certificate of Analysis

Research Use Only - Not for Use in Diagnostic Procedures

	Product	Material Number	Batch Number
Product Description:	AccuPlex™ SARS-CoV-2, Flu A/B and RSV Verification Panel v2	0505-0278	10576241

Date of Manufacture: 23-SEP-2021 Expiration Date: 23-SEP-2023

Storage Temperature: 2 - 8 °C. Average viral copies/mL for Member 1 determined using TaqMan® digital PCR Assays run on the BioRad QX200 system.

Kit Component	Analyte	Quantitative Result (cp/mL)
Member 1	SARS-CoV-2	9.44E+04
	Flu A	8.75E+04
	Flu B	9.15E+04
	RSV	9.90E+04


Qualitative detection using the Cepheid Xpert® Xpress.

QC Testing:

Kit Component	Result
Member 1	SARS-CoV-2 POSITIVE; Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE
Member 2	SARS-CoV-2 POSITIVE; Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE
Member 3	SARS-CoV-2 POSITIVE; Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE
Negative	SARS-CoV-2 NEGATIVE; Flu A NEGATIVE; Flu B NEGATIVE; RSV NEGATIVE

Safety Precautions: Preparation contains recombinant Alphavirus. Use the Centers for Disease Control recommended universal precautions.

Approval:

Prepared By:  Date: 17 NOV 2021

QA Verified By: _____ Date: 17 NOV 2021