Certificate of Analysis

Research Use Only - Not for Use in Diagnostic Procedures

Product Description:

Product	Material Number	Batch Number
AccuPlex™ SARS-CoV-2, Flu A/B and RSV Verification Panel v2	0505-0278	10799517

Date of Manufacture: 09 MAR 2026 **Expiration Date:** 09 MAR 2028

Storage Temperature: 2 - 8 °C. Average viral copies/mL for Member 1 determined using TaqMan® digital PCR Assays run on the BioRad QX200 system.

Kit Component	Analyte	Quantitative Result (cp/mL)
Member 1	SARS-CoV-2	26.7
	Flu A	26.6
	Flu B	25.3
	RSV	27.4

QC Testing:

Qualitative detection using the Cepheid Xpert® Xpress.

Kit Component	Result
Member 1	SARS-CoV-2 POSITIVE; Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE
Member 2	SARS-CoV-2 POSITIVE; Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE
Member 3	SARS-CoV-2 POSITIVE; Flu A POSITIVE; Flu B POSITIVE; RSV POSITIVE
Negative	SARS-CoV-2 NEGATIVE; Flu A NEGATIVE; Flu B NEGATIVE; RSV NEGATIVE

Safety Precautions: Preparation contains recombinant Alphavirus. Use the Centers for Disease Control recommended universal precautions.

Approval:

Prepared By — Date: 17 MAR 2026

QA Verified By — Date: 19 MAR 2026