

Certificate of Analysis Revised

Research Use Only - Not for Use in Diagnostic Procedures

	Product	Material Code	Batch Number
Product Description:	AccuPlex™ Flu A/B and RSV Reference Material Kit	0515-0001	10521218

Date of Manufacture: 22 SEP 2020

Expiration Date: 22 SEP 2024

Storage Temperature: 2 - 8 °C

Viral copies/mL as determined by Reverse Transcriptase digital PCR assay(s):

	Kit Component	Digital PCR target Sequence(s)	Result (cp/mL)
QC Testing:	Positive Reference Material	FluA, FluB and RSV genomic RNA	5343
	Negative Reference Material	Human RNaseP	3979

Cepheid Xpert® Xpress Flu/RSV Assay Testing

	Kit Component	Qualitative Result
QC Testing:	Positive Reference Material	Positive for FluA Positive for FluB Positive for RSV
	Negative Reference Material	Negative for FluA Negative for FluB Negative for RSV

Safety Precautions: Preparation contains recombinant Sindbis virus. Use the Centers for Disease Control recommended universal precautions.

Approval:


11 SEP 2023
 Prepared By Date


12 SEP 2023
 QA Verified By Date